

Test Report No. 7191246148-EEC20/03c-LDY
dated 20 Nov 2020

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.



PSB Singapore

**Add value.
Inspire trust.**

SUBJECT:

Testing of Gloves submitted by Lanhuo Medical Technology (Shanghai) Co., Ltd on 12 Oct 2020.

TESTED FOR:

Lanhuo Medical Technology (Shanghai) Co., Ltd
Room B, Second Floor,
Building 22, 189 Jinglian Road,
Minhang District, Shanghai, China

TEST DATE:

12 Oct 2020 to 07 Nov 2020

DESCRIPTION OF SAMPLES:

S/N	Product Description	Product Code	Brand/ Model	Colour	Lot No.	Size	Sample Received (pieces)	Manufacturer
1	Disposable Nitrile Multi-Purpose Gloves	LHEMB	(see Remark 1)	Blue	202000106598	L	408	Lanhuo Medical Technology (Jiangsu) Co., Ltd

Lot size as specified by client: 150,001 to 500,000 pieces

METHOD OF TEST:

EN 455-2:2015 Medical gloves for single use
Part 2: Requirements and testing for physical properties

- Clause 5 Strength
- Clause 7 Labelling



Laboratory:
TÜV SÜD PSB Pte. Ltd.
No.1 Science Park Drive
Singapore 118221

Phone : +65-6885 1333
Fax : +65-6776 8670
E-mail: enquiries@tuvsud.com
<https://www.tuvsud.com/en-sg>
Co. Reg : 199002667R

Regional Head Office:
TÜV SÜD Asia Pacific Pte. Ltd.
1 Science Park Drive, #02-01
Singapore 118221
TÜV®

Test Report No. 7191246148-EEC20/03c-LDY
dated 20 Nov 2020



PSB Singapore

RESULTS:

Sample: Disposable Nitrile Multi-Purpose Gloves, Blue, Size L

Table 1: Results for EN 455-2:2015 Clause 5

Clause	Tests	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
5	Strength a) Force at break (N)	For nitrile examination gloves: ≥ 6.0	13	7.6	Passed
	b) Force at break after challenge testing (N) 7 days at (70±2)°C	For nitrile examination gloves: ≥ 6.0	13	8.1	Passed

Table 2: Results for EN 455-2:2015 Clause 7

Clause	Tests	Requirements	Results	Inferred results
7	Labelling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012 and EN 1041:2008+A1:2013. Date of manufacture is defined as the packaging date.	Observed	Passed

REMARKS:

1. The Brand/Model was not provided by client.
2. Labelling requirements are assessed based on the submitted packaging artwork by client.

Yeo Poh Kwang
Associate Engineer

Lee Dai Yi
Engineer
Medical Health Services (NAM)

Test Report No. 7191246148-EEC20/03c-LDY
dated 20 Nov 2020



Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.
6. The tests carried out by TÜV SÜD PSB and this report are subject to TÜV SÜD PSB's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SÜD Group.

Effective 01 September 2020

